

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

AUG 13 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130117</u> ✓	
Name <u>KYLE VALERIO</u>		ID# <u>2093</u> ✓	Date <u>8/6/2010</u> ✓
A	Agency <u>KODIAK POLICE DEPARTMENT</u>		Phone # <u>(907)486-8000</u>
Instrument Location <u>217 LOWER MILL BAY ROAD, KODIAK, AK, 99615</u>			
Alco S/N <u>X301573</u> ✓		Target Value <u>.084</u> ✓	High Pressure <u>700</u>
B	Alco Test Values		
	<u>.078</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>			<u>COB</u> <u>8/16/10</u>
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>RD</sup> day of SEPT, 2010.

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

BT9 03/00



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

AUGUST 06, 2010

OPERATOR'S NAME:  
VALERIO, KYLE D  
OPERATOR'S NUMBER: 2093  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
TEST

D.L. #: XXXXXX  
DEPT/AGENCY: KOD1  
CASE/REPORT: 201000000  
TEST TYPE: V  
ALCO TARGET VALUE: .004  
ALCO S/N: X301573

— BREATH ANALYSIS —

.004 ADJUSTED FOR 29.88 in  
ALCO TARGET .003 13:49  
BLANK TEST .000 13:50  
INTERNAL STANDARD VERIFIED 13:50  
ALCO TV 29.88 in .078 13:50  
BLANK TEST .000 13:51  
SUBJECT SAMPLE .000 13:52  
BLANK TEST .000 13:52  
ALCO TV 29.88 in .079 13:53  
BLANK TEST .000 13:53

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

AUGUST 06, 2010  
TIME 13:53

— DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 50c  
BREATH TUBE: 41c  
BAROMETER: 29.98 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~